

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
 Data File : VN070718.D
 Acq On : 20 Jan 2022 14:21
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 21 02:13:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1043351	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1675629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1470752	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	512753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	700226	46.829	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.660%
35) Dibromofluoromethane	8.021	113	487515	46.904	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.800%
50) Toluene-d8	10.443	98	2061105	47.071	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.140%
62) 4-Bromofluorobenzene	12.731	95	657419	42.133	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.260%
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	133745	12.887	ug/l	99
52) Toluene	10.505	92	29341	0.905	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
Data File : VN070718.D
Acq On : 20 Jan 2022 14:21
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 21 02:13:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

