

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
 Data File : VN070719.D
 Acq On : 20 Jan 2022 14:47
 Operator : JC/MD
 Sample : PB142137TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB142137TB

Quant Time: Jan 21 02:14:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1010298	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1625774	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1489147	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	525275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	679089	46.901	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.800%
35) Dibromofluoromethane	8.024	113	470191	46.624	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.240%
50) Toluene-d8	10.440	98	2054780	48.366	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.740%
62) 4-Bromofluorobenzene	12.731	95	667370	44.083	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.160%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	73422	7.306	ug/l	96
20) Methylene Chloride	5.101	84	26077	1.970	ug/l	89
43) Isopropyl Acetate	8.563	43	108704	3.158	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
Data File : VN070719.D
Acq On : 20 Jan 2022 14:47
Operator : JC/MD
Sample : PB142137TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB142137TB

Quant Time: Jan 21 02:14:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

