

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
 Data File : VN070720.D
 Acq On : 20 Jan 2022 15:12
 Operator : JC/MD
 Sample : N1188-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11940

Quant Time: Jan 21 02:14:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1007744	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1620706	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1451498	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	499240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	684698	47.408	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.820%
35) Dibromofluoromethane	8.024	113	471160	46.866	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.740%
50) Toluene-d8	10.440	98	2052975	48.474	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.940%
62) 4-Bromofluorobenzene	12.731	95	645958	42.802	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.600%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	113429	11.316	ug/l	100
20) Methylene Chloride	5.098	84	20069	1.520	ug/l	90
43) Isopropyl Acetate	8.563	43	124103	3.616	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
Data File : VN070720.D
Acq On : 20 Jan 2022 15:12
Operator : JC/MD
Sample : N1188-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11940

Quant Time: Jan 21 02:14:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

