

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
 Data File : VN070722.D
 Acq On : 20 Jan 2022 16:12
 Operator : JC/MD
 Sample : N1195-01ME
 Misc : 5.09g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-011922ME

Quant Time: Jan 21 02:15:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/21/2022
 Supervised By : Mahesh Dadoda 01/21/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	951317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1545952	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1298691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	544108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	639007	46.869	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.740%
35) Dibromofluoromethane	8.019	113	434540	45.314	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.620%
50) Toluene-d8	10.441	98	1900989	47.056	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.120%
62) 4-Bromofluorobenzene	12.731	95	619701	43.048	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.100%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.854	43	28768m	1.546	ug/l	
25) 2-Butanone	7.348	43	105443	8.855	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
Data File : VN070722.D
Acq On : 20 Jan 2022 16:12
Operator : JC/MD
Sample : N1195-01ME
Misc : 5.09g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-02-011922ME

Quant Time: Jan 21 02:15:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/21/2022
Supervised By :Mahesh Dadoda 01/21/2022

