

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
 Data File : VN070710.D
 Acq On : 20 Jan 2022 10:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 21 02:09:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.578	0.549	5.0	90	0.00
3 P	Chloromethane	0.845	0.765	9.5	88	0.00
4 C	Vinyl Chloride	0.803	0.792	1.4#	93	0.00
5 T	Bromomethane	0.418	0.425	-1.7	95	0.00
6 T	Chloroethane	0.501	0.465	7.2	90	0.00
7 T	Trichlorofluoromethane	0.918	0.942	-2.6	97	0.00
8 T	Diethyl Ether	0.401	0.426	-6.2	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.555	0.519	6.5	91	0.00
10 T	Methyl Iodide	0.715	0.701	2.0	89	0.00
11 T	Tert butyl alcohol	0.129	0.125	3.1	93	0.00
12 CM	1,1-Dichloroethene	0.553	0.511	7.6#	91	0.00
13 T	Acrolein	0.185	0.254	-37.3#	137	0.00
14 T	Allyl chloride	1.205	1.196	0.7	90	0.00
15 T	Acrylonitrile	0.411	0.396	3.6	87	0.00
16 T	Acetone	0.497	0.501	-0.8	93	0.00
17 T	Carbon Disulfide	1.502	1.426	5.1	88	0.00
18 T	Methyl Acetate	0.978	0.898	8.2	88	0.00
19 T	Methyl tert-butyl Ether	1.976	1.992	-0.8	91	0.00
20 T	Methylene Chloride	0.655	0.626	4.4	91	0.00
21 T	trans-1,2-Dichloroethene	0.557	0.545	2.2	90	0.00
22 T	Diisopropyl ether	2.389	2.388	0.0	91	0.00
23 T	Vinyl Acetate	2.025	2.056	-1.5	90	0.00
24 P	1,1-Dichloroethane	1.195	1.169	2.2	90	0.00
25 T	2-Butanone	0.626	0.632	-1.0	90	0.00
26 T	2,2-Dichloropropane	0.916	0.936	-2.2	94	0.00
27 T	cis-1,2-Dichloroethene	0.661	0.645	2.4	90	0.00
28 T	Bromochloromethane	0.536	0.525	2.1	93	0.00
29 T	Tetrahydrofuran	0.383	0.371	3.1	86	0.00
30 C	Chloroform	1.342	1.117	16.8#	91	0.00
31 T	Cyclohexane	1.262	1.175	6.9	89	0.00
32 T	1,1,1-Trichloroethane	0.927	0.932	-0.5	92	0.00
33 S	1,2-Dichloroethane-d4	0.717	0.680	5.2	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.310	0.304	1.9	88	0.00
36 T	1,1-Dichloropropene	0.522	0.527	-1.0	92	0.00
37 T	Ethyl Acetate	0.693	0.672	3.0	86	0.00
38 T	Carbon Tetrachloride	0.464	0.476	-2.6	92	0.00
39 T	Methylcyclohexane	0.645	0.657	-1.9	90	0.00
40 TM	Benzene	1.618	1.598	1.2	90	0.00
41 T	Methacrylonitrile	0.353	0.352	0.3	91	0.00
42 TM	1,2-Dichloroethane	0.561	0.560	0.2	91	0.00
43 T	Isopropyl Acetate	1.059	1.075	-1.5	91	0.00
44 TM	Trichloroethene	0.367	0.361	1.6	89	0.00
45 C	1,2-Dichloropropane	0.440	0.441	-0.2#	91	0.00
46 T	Dibromomethane	0.256	0.258	-0.8	90	0.00
47 T	Bromodichloromethane	0.523	0.536	-2.5	93	0.00
48 T	Methyl methacrylate	0.504	0.528	-4.8	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
 Data File : VN070710.D
 Acq On : 20 Jan 2022 10:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 21 02:09:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	93	0.00
50 S	Toluene-d8	1.307	1.263	3.4	88	0.00
51 T	4-Methyl-2-Pentanone	0.676	0.688	-1.8	89	0.00
52 CM	Toluene	0.968	0.991	-2.4#	92	0.00
53 T	t-1,3-Dichloropropene	0.561	0.612	-9.1	94	0.00
54 T	cis-1,3-Dichloropropene	0.627	0.662	-5.6	92	0.00
55 T	1,1,2-Trichloroethane	0.380	0.379	0.3	91	0.00
56 T	Ethyl methacrylate	0.639	0.672	-5.2	91	0.00
57 T	1,3-Dichloropropane	0.673	0.683	-1.5	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.151	0.173	-14.6	95	0.00
59 T	2-Hexanone	0.491	0.534	-8.8	91	0.00
60 T	Dibromochloromethane	0.355	0.380	-7.0	93	0.00
61 T	1,2-Dibromoethane	0.382	0.384	-0.5	90	0.00
62 S	4-Bromofluorobenzene	0.466	0.468	-0.4	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.367	0.353	3.8	90	0.00
65 PM	Chlorobenzene	1.083	1.081	0.2	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.381	-2.1	92	0.00
67 C	Ethyl Benzene	2.008	2.065	-2.8#	92	0.00
68 T	m/p-Xylenes	0.731	0.754	-3.1	91	0.00
69 T	o-Xylene	0.725	0.751	-3.6	92	0.00
70 T	Styrene	1.160	1.239	-6.8	92	0.00
71 P	Bromoform	0.282	0.302	-7.1	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.730	4.675	1.2	92	0.00
74 T	N-amyl acetate	1.934	2.012	-4.0	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.598	1.500	6.1	88	0.00
76 T	1,2,3-Trichloropropane	1.326	1.301	1.9	96	0.00
77 T	Bromobenzene	1.061	1.057	0.4	92	0.00
78 T	n-propylbenzene	5.290	5.455	-3.1	92	0.00
79 T	2-Chlorotoluene	3.346	3.325	0.6	92	0.00
80 T	1,3,5-Trimethylbenzene	3.825	3.906	-2.1	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.445	0.468	-5.2	94	0.00
82 T	4-Chlorotoluene	3.127	3.250	-3.9	93	0.00
83 T	tert-Butylbenzene	3.330	3.316	0.4	91	0.00
84 T	1,2,4-Trimethylbenzene	3.677	3.809	-3.6	92	0.00
85 T	sec-Butylbenzene	4.527	4.693	-3.7	92	0.00
86 T	p-Isopropyltoluene	3.556	3.770	-6.0	92	0.00
87 T	1,3-Dichlorobenzene	1.764	1.811	-2.7	91	0.00
88 T	1,4-Dichlorobenzene	1.756	1.760	-0.2	92	0.00
89 T	n-Butylbenzene	2.934	3.175	-8.2	94	0.00
90 T	Hexachloroethane	0.622	0.637	-2.4	92	0.00
91 T	1,2-Dichlorobenzene	1.743	1.725	1.0	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.271	0.264	2.6	88	0.00
93 T	1,2,4-Trichlorobenzene	0.774	0.851	-9.9	90	0.00
94 T	Hexachlorobutadiene	0.516	0.539	-4.5	94	0.00
95 T	Naphthalene	2.001	2.232	-11.5	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
Data File : VN070710.D
Acq On : 20 Jan 2022 10:01
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 21 02:09:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.738	0.824	-11.7	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6