

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012023\
 Data File : VN076338.D
 Acq On : 20 Jan 2023 11:01
 Operator : JC\MD
 Sample : VN0120MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0120MBL01

Quant Time: Jan 20 22:11:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	356885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	563440	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	473770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	187829	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	188929	47.280	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.560%
35) Dibromofluoromethane	8.171	113	168437	47.790	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.580%
50) Toluene-d8	10.571	98	636557	48.904	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.800%
62) 4-Bromofluorobenzene	12.853	95	195749	46.066	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012023\
Data File : VN076338.D
Acq On : 20 Jan 2023 11:01
Operator : JC\MD
Sample : VN0120MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0120MBL01

Quant Time: Jan 20 22:11:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

