

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012023\
 Data File : VN076347.D
 Acq On : 20 Jan 2023 14:49
 Operator : JC\MD
 Sample : 01233-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB01

Quant Time: Jan 20 22:16:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	370105	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	595933	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	523917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	212240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	203331	49.066	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.140%
35) Dibromofluoromethane	8.177	113	179169	48.063	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	10.571	98	682616	49.583	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.160%
62) 4-Bromofluorobenzene	12.853	95	215515	47.952	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.900%
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.295	84	13253	2.803	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012023\
Data File : VN076347.D
Acq On : 20 Jan 2023 14:49
Operator : JC\MD
Sample : 01233-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB01

Quant Time: Jan 20 22:16:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

