

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012023\
 Data File : VN076348.D
 Acq On : 20 Jan 2023 15:26
 Operator : JC\MD
 Sample : 01233-12RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB01RE

Quant Time: Jan 20 22:16:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	382643	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	610377	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	521259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	195741	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	206927	48.298	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.600%
35) Dibromofluoromethane	8.177	113	184326	48.276	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.560%
50) Toluene-d8	10.571	98	694768	49.271	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.540%
62) 4-Bromofluorobenzene	12.853	95	208192	45.227	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.460%
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	5325	3.805	ug/l	93
20) Methylene Chloride	5.289	84	11468	2.246	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012023\
Data File : VN076348.D
Acq On : 20 Jan 2023 15:26
Operator : JC\MD
Sample : 01233-12RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB01RE

Quant Time: Jan 20 22:16:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

