

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012119\
 Data File : VN053495.D
 Acq On : 21 Jan 2019 10:06
 Operator : JC/SP
 Sample : VN0121WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0121WBL01

Quant Time: Jan 22 04:26:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	833469	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1297368	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1214983	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	557351	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	408209	49.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.10%	
35) Dibromofluoromethane	7.59	113	391928	52.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.86%	
50) Toluene-d8	10.09	98	1608669	51.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.74%	
62) 4-Bromofluorobenzene	12.40	95	548995	50.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012119\
Data File : VN053495.D
Acq On : 21 Jan 2019 10:06
Operator : JC/SP
Sample : VN0121WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN0121WBL01

Quant Time: Jan 22 04:26:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

