

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012120\
 Data File : VN059810.D
 Acq On : 21 Jan 2020 14:31
 Operator : JC/MD
 Sample : PB126227ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126227ZHE#11

Quant Time: Jan 22 03:31:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	202196	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	304293	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	272337	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	112611	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	107842	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
35) Dibromofluoromethane	7.57	113	93325	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
50) Toluene-d8	10.08	98	361145	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
62) 4-Bromofluorobenzene	12.40	95	120781	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
Target Compounds						
20) Methylene Chloride	4.53	84	7369	2.395	ug/l	90
43) Isopropyl Acetate	8.14	43	31835	7.447	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012120\
Data File : VN059810.D
Acq On : 21 Jan 2020 14:31
Operator : JC/MD
Sample : PB126227ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126227ZHE#11

Quant Time: Jan 22 03:31:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

