

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012122\
 Data File : VN070732.D
 Acq On : 21 Jan 2022 15:35
 Operator : JC/MD
 Sample : N1217-01 200X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2250

Quant Time: Jan 22 05:29:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	962614	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1526936	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.741	117	1342618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	498605	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	649108	47.051	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.100%
35) Dibromofluoromethane	8.021	113	443328	46.806	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.620%
50) Toluene-d8	10.438	98	1881753	47.160	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.320%
62) 4-Bromofluorobenzene	12.728	95	618099	43.471	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.940%
Target Compounds						
					Qvalue	
16) Acetone	4.285	43	523713	54.694	ug/l	97
52) Toluene	10.502	92	94919	3.212	ug/l	99
84) 1,2,4-Trimethylbenzene	13.366	105	43818	1.195	ug/l	93
93) 1,2,4-Trichlorobenzene	15.257	180	9881	4.424	ug/l #	85
95) Naphthalene	15.504	128	63506	7.162	ug/l	98
96) 1,2,3-Trichlorobenzene	15.694	180	11082	4.394	ug/l #	20

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012122\
Data File : VN070732.D
Acq On : 21 Jan 2022 15:35
Operator : JC/MD
Sample : N1217-01 200X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2250

Quant Time: Jan 22 05:29:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

