

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012122\
 Data File : VN070736.D
 Acq On : 21 Jan 2022 17:16
 Operator : JC/MD
 Sample : PB142170ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB142170ZHE#03

Quant Time: Jan 22 05:30:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1037356	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1685666	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.741	117	1496850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	547570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	703957	47.351	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.700%
35) Dibromofluoromethane	8.021	113	486883	46.564	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.120%
50) Toluene-d8	10.438	98	2102546	47.732	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.460%
62) 4-Bromofluorobenzene	12.728	95	690472	43.988	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.980%
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	81082	7.858	ug/l	97
20) Methylene Chloride	5.098	84	19139	1.408	ug/l	90
43) Isopropyl Acetate	8.560	43	132191	3.703	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012122\
Data File : VN070736.D
Acq On : 21 Jan 2022 17:16
Operator : JC/MD
Sample : PB142170ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB142170ZHE#03

Quant Time: Jan 22 05:30:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

