

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012122\
 Data File : VN070742.D
 Acq On : 21 Jan 2022 19:48
 Operator : JC/MD
 Sample : N1217-08
 Misc : 1.07g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-242

Quant Time: Jan 22 05:31:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	909648	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	1480738	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1290930	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	559088	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.429	65	621291	47.657	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.320%
35) Dibromofluoromethane	8.016	113	423109	46.065	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.140%
50) Toluene-d8	10.438	98	1836906	47.472	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.940%
62) 4-Bromofluorobenzene	12.731	95	627729	45.526	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.060%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.862	43	29334	1.648	ug/l	# 72
25) 2-Butanone	7.343	43	88685	7.789	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012122\
Data File : VN070742.D
Acq On : 21 Jan 2022 19:48
Operator : JC/MD
Sample : N1217-08
Misc : 1.07g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-242

Quant Time: Jan 22 05:31:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

