

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012125\
 Data File : VN085514.D
 Acq On : 21 Jan 2025 16:08
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 22 00:41:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	109	0.00
2 T	Dichlorodifluoromethane	0.677	0.644	4.9	111	0.00
3 P	Chloromethane	0.733	0.625	14.7	100	0.00
4 C	Vinyl Chloride	0.737	0.676	8.3#	107	0.00
5 T	Bromomethane	0.445	0.404	9.2	105	0.00
6 T	Chloroethane	0.467	0.439	6.0	113	0.00
7 T	Trichlorofluoromethane	1.069	1.079	-0.9	118	0.00
8 T	Diethyl Ether	0.369	0.331	10.3	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.602	0.593	1.5	119	0.00
10 T	Methyl Iodide	0.690	0.683	1.0	103	0.00
11 T	Tert butyl alcohol	0.092	0.067	27.2#	79	-0.01
12 CM	1,1-Dichloroethene	0.537	0.526	2.0#	107	0.00
13 T	Acrolein	0.126	0.110	12.7	86	0.00
14 T	Allyl chloride	0.871	0.843	3.2	108	0.00
15 T	Acrylonitrile	0.294	0.247	16.0	90	0.00
16 T	Acetone	0.261	0.195	25.3#	84	0.00
17 T	Carbon Disulfide	1.652	1.439	12.9	106	0.00
18 T	Methyl Acetate	0.793	0.685	13.6	94	0.00
19 T	Methyl tert-butyl Ether	1.742	1.763	-1.2	102	0.00
20 T	Methylene Chloride	0.646	0.605	6.3	105	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.560	2.3	110	0.00
22 T	Diisopropyl ether	1.933	1.968	-1.8	105	0.00
23 T	Vinyl Acetate	1.353	1.356	-0.2	99	0.00
24 P	1,1-Dichloroethane	1.178	1.145	2.8	106	0.00
25 T	2-Butanone	0.384	0.310	19.3	87	0.00
26 T	2,2-Dichloropropane	0.953	1.045	-9.7	122	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.679	-0.6	108	0.00
28 T	Bromochloromethane	0.548	0.461	15.9	93	0.00
29 T	Tetrahydrofuran	0.243	0.205	15.6	86	0.00
30 C	Chloroform	1.218	1.179	3.2#	109	0.00
31 T	Cyclohexane	1.024	0.932	9.0	115	0.00
32 T	1,1,1-Trichloroethane	1.068	1.052	1.5	113	0.00
33 S	1,2-Dichloroethane-d4	0.807	0.694	14.0	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35 S	Dibromofluoromethane	0.347	0.314	9.5	92	0.00
36 T	1,1-Dichloropropene	0.487	0.496	-1.8	113	0.00
37 T	Ethyl Acetate	0.491	0.394	19.8	85	0.00
38 T	Carbon Tetrachloride	0.557	0.563	-1.1	111	0.00
39 T	Methylcyclohexane	0.457	0.518	-13.3	118	0.00
40 TM	Benzene	1.463	1.489	-1.8	108	0.00
41 T	Methacrylonitrile	0.256	0.241	5.9	94	0.00
42 TM	1,2-Dichloroethane	0.551	0.544	1.3	104	0.00
43 T	Isopropyl Acetate	0.787	0.741	5.8	96	0.00
44 TM	Trichloroethene	0.341	0.343	-0.6	111	0.00
45 C	1,2-Dichloropropane	0.374	0.371	0.8#	105	0.00
46 T	Dibromomethane	0.270	0.255	5.6	100	0.00
47 T	Bromodichloromethane	0.549	0.574	-4.6	108	0.00
48 T	Methyl methacrylate	0.354	0.352	0.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012125\
 Data File : VN085514.D
 Acq On : 21 Jan 2025 16:08
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 22 00:41:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.004	33.3#	71	0.00
50 S	Toluene-d8	1.232	1.164	5.5	96	0.00
51 T	4-Methyl-2-Pentanone	0.457	0.416	9.0	89	0.00
52 CM	Toluene	0.848	0.913	-7.7#	110	0.00
53 T	t-1,3-Dichloropropene	0.519	0.559	-7.7	106	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.606	-9.4	108	0.00
55 T	1,1,2-Trichloroethane	0.335	0.335	0.0	103	0.00
56 T	Ethyl methacrylate	0.469	0.509	-8.5	98	0.00
57 T	1,3-Dichloropropane	0.583	0.587	-0.7	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.199	6.6	87	0.00
59 T	2-Hexanone	0.321	0.288	10.3	85	0.00
60 T	Dibromochloromethane	0.405	0.416	-2.7	105	0.00
61 T	1,2-Dibromoethane	0.334	0.328	1.8	103	0.00
62 S	4-Bromofluorobenzene	0.422	0.411	2.6	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.341	0.364	-6.7	121	0.00
65 PM	Chlorobenzene	1.095	1.103	-0.7	110	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.399	0.7	109	0.00
67 C	Ethyl Benzene	1.784	1.984	-11.2#	114	0.00
68 T	m/p-Xylenes	0.659	0.743	-12.7	112	0.00
69 T	o-Xylene	0.630	0.704	-11.7	111	0.00
70 T	Styrene	1.043	1.209	-15.9	110	0.00
71 P	Bromoform	0.288	0.287	0.3	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	3.375	3.802	-12.7	115	0.00
74 T	N-amyl acetate	1.517	1.449	4.5	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.192	1.033	13.3	94	0.00
76 T	1,2,3-Trichloropropane	1.016	0.909	10.5	106	0.00
77 T	Bromobenzene	0.882	0.893	-1.2	109	0.00
78 T	n-propylbenzene	3.995	4.582	-14.7	115	0.00
79 T	2-Chlorotoluene	2.586	2.734	-5.7	110	0.00
80 T	1,3,5-Trimethylbenzene	2.787	3.230	-15.9	115	0.00
81 T	trans-1,4-Dichloro-2-butene	0.376	0.367	2.4	98	0.00
82 T	4-Chlorotoluene	2.574	2.813	-9.3	113	0.00
83 T	tert-Butylbenzene	2.341	2.724	-16.4	117	0.00
84 T	1,2,4-Trimethylbenzene	2.778	3.226	-16.1	112	0.00
85 T	sec-Butylbenzene	3.244	3.828	-18.0	117	0.00
86 T	p-Isopropyltoluene	2.631	3.214	-22.2	117	0.00
87 T	1,3-Dichlorobenzene	1.624	1.669	-2.8	111	0.00
88 T	1,4-Dichlorobenzene	1.683	1.674	0.5	112	0.00
89 T	n-Butylbenzene	2.327	2.801	-20.4	118	0.00
90 T	Hexachloroethane	0.618	0.604	2.3	109	0.00
91 T	1,2-Dichlorobenzene	1.620	1.576	2.7	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.218	0.181	17.0	85	0.00
93 T	1,2,4-Trichlorobenzene	0.753	0.791	-5.0	106	0.00
94 T	Hexachlorobutadiene	0.400	0.391	2.3	111	0.00
95 T	Naphthalene	2.246	2.113	5.9	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012125\
Data File : VN085514.D
Acq On : 21 Jan 2025 16:08
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 22 00:41:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.762	0.743	2.5	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6