


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012219\
Data File : VN053507.D
Acq On : 22 Jan 2019 8:35
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampled :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Title : SW846 8260
Last Update : Wed Jan 16 03:10:53 2019



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	14513	PASS
75	95	30	60	47.1	42408	PASS
95	95	100	100	100.0	90053	PASS
96	95	5	9	6.6	5948	PASS
173	174	0.00	2	0.9	780	PASS
174	95	50	100	95.1	85664	PASS
175	174	5	9	7.2	6144	PASS
176	174	95	101	96.2	82448	PASS
177	176	5	9	6.4	5285	PASS