

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012224\
 Data File : VN080746.D
 Acq On : 22 Jan 2024 19:21
 Operator : JC\MD
 Sample : P1179-13 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-6

Quant Time: Jan 23 03:58:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	124509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	296628	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	302150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122070	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	100788	46.095	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.200%		
35) Dibromofluoromethane	8.165	113	84045	49.867	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.740%		
50) Toluene-d8	10.565	98	338822	52.723	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	105.440%		
62) 4-Bromofluorobenzene	12.853	95	133909	54.338	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	108.680%		
Target Compounds						
					Qvalue	
40) Benzene	8.606	78	395273	40.728	ug/l	100
52) Toluene	10.629	92	131437	23.637	ug/l	100
68) m/p-Xylenes	12.071	106	20453	4.615	ug/l	85
69) o-Xylene	12.400	106	12831	2.961	ug/l	93
95) Naphthalene	15.641	128	70673	8.309	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012224\
Data File : VN080746.D
Acq On : 22 Jan 2024 19:21
Operator : JC\MD
Sample : P1179-13 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-6

Quant Time: Jan 23 03:58:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

