

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012224\
 Data File : VN080747.D
 Acq On : 22 Jan 2024 19:45
 Operator : JC\MD
 Sample : P1179-14 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-7

Quant Time: Jan 23 03:59:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	124865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	299203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	313062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	129013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	104512	47.662	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.320%
35) Dibromofluoromethane	8.165	113	83651	49.207	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.420%
50) Toluene-d8	10.565	98	346721	53.488	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.980%
62) 4-Bromofluorobenzene	12.847	95	142518	57.333	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.660%
Target Compounds						
					Qvalue	
40) Benzene	8.606	78	143751	14.684	ug/l	97
52) Toluene	10.629	92	58990	10.517	ug/l	98
68) m/p-Xylenes	12.065	106	13137	2.861	ug/l	93
69) o-Xylene	12.406	106	10174	2.266	ug/l	99
95) Naphthalene	15.641	128	56766	6.315	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012224\
Data File : VN080747.D
Acq On : 22 Jan 2024 19:45
Operator : JC\MD
Sample : P1179-14 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-7

Quant Time: Jan 23 03:59:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

