

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012224\
 Data File : VN080748.D
 Acq On : 22 Jan 2024 20:08
 Operator : JC\MD
 Sample : P1179-15 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-8

Quant Time: Jan 23 03:59:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.223	168	129229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302871	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	310195	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	126015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	99346	43.776	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.560%
35) Dibromofluoromethane	8.165	113	83085	48.282	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.560%
50) Toluene-d8	10.564	98	345252	52.617	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.240%
62) 4-Bromofluorobenzene	12.847	95	133239	52.951	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.900%
Target Compounds						
					Qvalue	
40) Benzene	8.606	78	170735	17.229	ug/l	98
52) Toluene	10.629	92	79428	13.990	ug/l	99
68) m/p-Xylenes	12.070	106	19139	4.207	ug/l	97
69) o-Xylene	12.400	106	13626	3.063	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	10901	1.110	ug/l	96
95) Naphthalene	15.641	128	77430	8.819	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012224\
Data File : VN080748.D
Acq On : 22 Jan 2024 20:08
Operator : JC\MD
Sample : P1179-15 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-8

Quant Time: Jan 23 03:59:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

