

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012323\
 Data File : VN076361.D
 Acq On : 23 Jan 2023 13:00
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 23 13:18:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	371785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	596895	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	504710	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	200062	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	208895	50.181	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.360%
35) Dibromofluoromethane	8.171	113	182799	48.958	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.920%
50) Toluene-d8	10.571	98	673279	48.826	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.660%
62) 4-Bromofluorobenzene	12.853	95	207242	46.037	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012323\
Data File : VN076361.D
Acq On : 23 Jan 2023 13:00
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 23 13:18:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

