

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012324\
 Data File : VN080754.D
 Acq On : 23 Jan 2024 10:55
 Operator : JC\MD
 Sample : VN0123WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0123WBL01

Quant Time: Jan 24 01:54:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	125743	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	295558	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122356	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	97346	44.084	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.160%
35) Dibromofluoromethane	8.165	113	82892	49.361	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.720%
50) Toluene-d8	10.565	98	341575	53.344	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.680%
62) 4-Bromofluorobenzene	12.847	95	133770	54.478	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012324\
Data File : VN080754.D
Acq On : 23 Jan 2024 10:55
Operator : JC\MD
Sample : VN0123WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0123WBL01

Quant Time: Jan 24 01:54:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

