

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012324\
 Data File : VN080758.D
 Acq On : 23 Jan 2024 13:14
 Operator : JC\MD
 Sample : P1179-11DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-4DL

Quant Time: Jan 24 01:58:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	122439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	285216	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	300988	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	124519	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	94683	44.035	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	88.080%		
35) Dibromofluoromethane	8.165	113	79202	48.874	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.740%		
50) Toluene-d8	10.565	98	330005	53.406	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	106.820%		
62) 4-Bromofluorobenzene	12.847	95	131529	55.507	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	111.020%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	4821	6.829	ug/l #	59
20) Methylene Chloride	5.271	84	2136	1.037	ug/l #	80
40) Benzene	8.606	78	225920	24.209	ug/l	100
43) Isopropyl Acetate	8.682	43	27926	4.720	ug/l #	83
52) Toluene	10.629	92	140625	26.301	ug/l	98
67) Ethyl Benzene	11.965	91	19560	1.558	ug/l	96
68) m/p-Xylenes	12.065	106	37273	8.443	ug/l	89
69) o-Xylene	12.400	106	27556	6.384	ug/l	82
84) 1,2,4-Trimethylbenzene	13.482	105	23496	2.421	ug/l	97
95) Naphthalene	15.641	128	259999	29.967	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012324\
Data File : VN080758.D
Acq On : 23 Jan 2024 13:14
Operator : JC\MD
Sample : P1179-11DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-4DL

Quant Time: Jan 24 01:58:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

