

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012324\
 Data File : VN080759.D
 Acq On : 23 Jan 2024 13:48
 Operator : JC\MD
 Sample : P1183-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-305-0-2

Quant Time: Jan 24 01:59:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	117328	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	280951	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	296274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	121729	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	97345	47.246	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.500%		
35) Dibromofluoromethane	8.165	113	75692	47.417	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.840%		
50) Toluene-d8	10.571	98	321475	52.815	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	105.640%		
62) 4-Bromofluorobenzene	12.853	95	126157	54.049	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	108.100%		
Target Compounds						
16) Acetone	4.430	43	17560	25.959	ug/l	95
20) Methylene Chloride	5.277	84	6982	3.536	ug/l #	86
25) 2-Butanone	7.483	43	9466	7.742	ug/l #	76
43) Isopropyl Acetate	8.688	43	117960	20.239	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012324\
Data File : VN080759.D
Acq On : 23 Jan 2024 13:48
Operator : JC\MD
Sample : P1183-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-305-0-2

Quant Time: Jan 24 01:59:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

