

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012324\
 Data File : VN080760.D
 Acq On : 23 Jan 2024 14:12
 Operator : JC\MD
 Sample : P1183-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-305-25-27

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/24/2024
 Supervised By : Mahesh Dadoda 01/24/2024

Quant Time: Jan 24 02:00:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	119136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	271239	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	298458	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122287	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	93635	44.755	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.520%		
35) Dibromofluoromethane	8.171	113	75874	49.233	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.460%		
50) Toluene-d8	10.565	98	325110	55.325	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	110.660%		
62) 4-Bromofluorobenzene	12.847	95	129251	57.357	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	114.720%		
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	22719	33.076	ug/l #	85
20) Methylene Chloride	5.283	84	16393	8.175	ug/l #	80
25) 2-Butanone	7.488	43	10599m	8.537	ug/l	
43) Isopropyl Acetate	8.688	43	118384	21.039	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012324\
Data File : VN080760.D
Acq On : 23 Jan 2024 14:12
Operator : JC\MD
Sample : P1183-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-305-25-27

Quant Time: Jan 24 02:00:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/24/2024
Supervised By :Mahesh Dadoda 01/24/2024

