

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012420\
 Data File : VN059835.D
 Acq On : 24 Jan 2020 10:48
 Operator : JC/MD
 Sample : VN0124WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0124WBL01

Quant Time: Jan 25 00:50:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	209157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	313014	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	267567	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	104099	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	110089	47.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.80%	
35) Dibromofluoromethane	7.56	113	95404	51.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.22%	
50) Toluene-d8	10.08	98	363212	50.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.44%	
62) 4-Bromofluorobenzene	12.40	95	116194	44.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012420\
Data File : VN059835.D
Acq On : 24 Jan 2020 10:48
Operator : JC/MD
Sample : VN0124WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0124WBL01

Quant Time: Jan 25 00:50:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

