

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012420\
 Data File : VN059838.D
 Acq On : 24 Jan 2020 12:23
 Operator : JC/MD
 Sample : L1215-01ME
 Misc : 4.17g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RBR-200055ME

Quant Time: Jan 25 00:56:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	198150	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	294098	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	261004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	122841	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	103414	47.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.00%	
35) Dibromofluoromethane	7.56	113	87596	49.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.90%	
50) Toluene-d8	10.08	98	347965	51.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.42%	
62) 4-Bromofluorobenzene	12.40	95	119224	48.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.34%	
Target Compounds						
52) Toluene	10.14	92	5412	1.122	ug/l	99
67) Ethyl Benzene	11.50	91	20321	2.182	ug/l	98
68) m/p-Xylenes	11.61	106	35277	10.028	ug/l	94
69) o-Xylene	11.94	106	11000	3.194	ug/l	94
78) n-propylbenzene	12.59	91	16844	1.666	ug/l	97
80) 1,3,5-Trimethylbenzene	12.73	105	25631	3.383	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	90264	12.042	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012420\
Data File : VN059838.D
Acq On : 24 Jan 2020 12:23
Operator : JC/MD
Sample : L1215-01ME
Misc : 4.17µ/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RBR-200055ME

Quant Time: Jan 25 00:56:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

