

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012422\
 Data File : VN070754.D
 Acq On : 24 Jan 2022 14:38
 Operator : JC/MD
 Sample : PB142209ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB142209ZHE#02

Quant Time: Jan 24 23:31:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1049677	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1683071	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1501055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	542266	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	701560	46.635	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.280%
35) Dibromofluoromethane	8.021	113	487363	46.682	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.360%
50) Toluene-d8	10.440	98	2082268	47.344	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.680%
62) 4-Bromofluorobenzene	12.731	95	686541	43.805	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.620%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	90459	8.664	ug/l	99
20) Methylene Chloride	5.098	84	26163	1.902	ug/l #	86
43) Isopropyl Acetate	8.560	43	123947	3.478	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012422\
Data File : VN070754.D
Acq On : 24 Jan 2022 14:38
Operator : JC/MD
Sample : PB142209ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB142209ZHE#02

Quant Time: Jan 24 23:31:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

