

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012422\
 Data File : VN070755.D
 Acq On : 24 Jan 2022 15:04
 Operator : JC/MD
 Sample : PB142209ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB142209ZHE#03

Quant Time: Jan 24 23:32:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1007436	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1628990	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1459694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	544163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	678712	47.008	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.020%		
35) Dibromofluoromethane	8.021	113	468489	46.364	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.720%		
50) Toluene-d8	10.440	98	2026841	47.614	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.220%		
62) 4-Bromofluorobenzene	12.731	95	669286	44.122	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.240%		
Target Compounds						
16) Acetone	4.291	43	90812	9.062	ug/l	98
20) Methylene Chloride	5.093	84	26014	1.971	ug/l	93
43) Isopropyl Acetate	8.560	43	134400	3.896	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012422\
Data File : VN070755.D
Acq On : 24 Jan 2022 15:04
Operator : JC/MD
Sample : PB142209ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB142209ZHE#03

Quant Time: Jan 24 23:32:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

