

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012422\
 Data File : VN070758.D
 Acq On : 24 Jan 2022 16:20
 Operator : JC/MD
 Sample : N1224-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AREA-1-B-2@4-BGS

Quant Time: Jan 24 23:33:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1002233	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1627653	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1453975	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	538415	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	684198	47.634	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.260%
35) Dibromofluoromethane	8.021	113	467712	46.325	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.640%
50) Toluene-d8	10.440	98	2016182	47.402	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.800%
62) 4-Bromofluorobenzene	12.731	95	665514	43.909	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.820%
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	167788	16.830	ug/l	99
18) Methyl Acetate	4.865	43	28322	1.445	ug/l #	81
20) Methylene Chloride	5.095	84	46363	3.531	ug/l	89
43) Isopropyl Acetate	8.563	43	139037	4.034	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012422\
Data File : VN070758.D
Acq On : 24 Jan 2022 16:20
Operator : JC/MD
Sample : N1224-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AREA-1-B-2@4-BGS

Quant Time: Jan 24 23:33:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

