

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012422\
 Data File : VN070761.D
 Acq On : 24 Jan 2022 17:36
 Operator : JC/MD
 Sample : N1235-04
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Jan 24 23:33:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	926936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1501787	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1232424	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	515337	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	629894	47.416	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.840%
35) Dibromofluoromethane	8.016	113	424778	45.599	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.200%
50) Toluene-d8	10.440	98	1815724	46.267	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	92.540%
62) 4-Bromofluorobenzene	12.731	95	584617	41.805	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.600%
Target Compounds						
18) Methyl Acetate	4.862	43	83932	4.629	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012422\
Data File : VN070761.D
Acq On : 24 Jan 2022 17:36
Operator : JC/MD
Sample : N1235-04
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Jan 24 23:33:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

