

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012422\
 Data File : VN070763.D
 Acq On : 24 Jan 2022 18:27
 Operator : JC/MD
 Sample : N1242-03ME
 Misc : 3.68g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PSC-124055ME

Quant Time: Jan 24 23:34:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	926108	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1504260	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1253707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	507670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	631945	47.613	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.220%
35) Dibromofluoromethane	8.016	113	418932	44.897	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.800%
50) Toluene-d8	10.440	98	1843019	46.885	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.780%
62) 4-Bromofluorobenzene	12.731	95	592624	42.308	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.620%
Target Compounds						Qvalue
18) Methyl Acetate	4.862	43	56540	3.121	ug/l	# 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012422\
Data File : VN070763.D
Acq On : 24 Jan 2022 18:27
Operator : JC/MD
Sample : N1242-03ME
Misc : 3.68g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PSC-124055ME

Quant Time: Jan 24 23:34:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

