

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012423\
 Data File : VN076380.D
 Acq On : 24 Jan 2023 14:17
 Operator : JC\MD
 Sample : 01280-04DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 22-MW-3S

Quant Time: Jan 25 00:39:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	403734	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	640568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	547089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	220043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	227118	50.241	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.480%
35) Dibromofluoromethane	8.171	113	196087	48.936	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	10.571	98	739035	49.940	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.880%
62) 4-Bromofluorobenzene	12.853	95	227406	47.072	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.140%
Target Compounds						
						Qvalue
95) Naphthalene	15.647	128	3021	2.093	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012423\
Data File : VN076380.D
Acq On : 24 Jan 2023 14:17
Operator : JC\MD
Sample : 01280-04DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
22-MW-3S

Quant Time: Jan 25 00:39:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

