

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012423\
 Data File : VN076381.D
 Acq On : 24 Jan 2023 14:41
 Operator : JC\MD
 Sample : 01280-02 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 22-MW-4I

Quant Time: Jan 25 00:39:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	383856	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	620713	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	530925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	222241	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	217675	50.646	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.300%
35) Dibromofluoromethane	8.171	113	186238	47.965	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.920%
50) Toluene-d8	10.570	98	711088	49.589	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.180%
62) 4-Bromofluorobenzene	12.853	95	224007	47.852	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.700%
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	7.494	96	28524	6.140	ug/l	93
44) Trichloroethene	9.359	130	121934	26.132	ug/l	96
95) Naphthalene	15.647	128	1808	1.990	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012423\
Data File : VN076381.D
Acq On : 24 Jan 2023 14:41
Operator : JC\MD
Sample : 01280-02 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
22-MW-4I

Quant Time: Jan 25 00:39:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

