

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012423\
 Data File : VN076382.D
 Acq On : 24 Jan 2023 15:05
 Operator : JC\MD
 Sample : 01280-01 400X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LEA-7I

Quant Time: Jan 25 00:39:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	400603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	635764	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	546837	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	217371	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	221394	49.358	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.720%		
35) Dibromofluoromethane	8.177	113	189858	47.739	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.480%		
50) Toluene-d8	10.571	98	730864	49.761	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.520%		
62) 4-Bromofluorobenzene	12.853	95	229087	47.779	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.560%		
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.500	96	7836	1.616	ug/l	93
44) Trichloroethene	9.359	130	141069	29.517	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012423\
Data File : VN076382.D
Acq On : 24 Jan 2023 15:05
Operator : JC\MD
Sample : 01280-01 400X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LEA-7I

Quant Time: Jan 25 00:39:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

