

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012423\
 Data File : VN076385.D
 Acq On : 24 Jan 2023 16:17
 Operator : JC\MD
 Sample : 01280-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 22-MW-3I

Quant Time: Jan 25 00:40:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	369456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	597442	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	514256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	200889	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	208578	50.421	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.840%
35) Dibromofluoromethane	8.171	113	182625	48.866	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.740%
50) Toluene-d8	10.571	98	684962	49.627	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.260%
62) 4-Bromofluorobenzene	12.853	95	210760	46.776	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.560%
Target Compounds						
						Qvalue
44) Trichloroethene	9.359	130	9643	2.147	ug/l	92
95) Naphthalene	15.635	128	940	1.926	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012423\
Data File : VN076385.D
Acq On : 24 Jan 2023 16:17
Operator : JC\MD
Sample : 01280-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
22-MW-3I

Quant Time: Jan 25 00:40:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

