

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012423\
 Data File : VN076387.D
 Acq On : 24 Jan 2023 17:58
 Operator : JC\MD
 Sample : 01285-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2023-1-24

Quant Time: Jan 25 00:41:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	364837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	588948	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	509956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	201122	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	210447	51.517	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.040%
35) Dibromofluoromethane	8.171	113	179205	48.643	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.280%
50) Toluene-d8	10.571	98	680678	50.028	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.060%
62) 4-Bromofluorobenzene	12.847	95	211422	47.599	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012423\
Data File : VN076387.D
Acq On : 24 Jan 2023 17:58
Operator : JC\MD
Sample : 01285-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2023-1-24

Quant Time: Jan 25 00:41:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

