

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
 Data File : VN065574.D
 Acq On : 25 Jan 2021 13:13
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001

Quant Time: Jan 25 13:39:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:39:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	166385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	301352	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	255297	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	93443	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	3120	1.43	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	2.86%#
35) Dibromofluoromethane	7.543	113	1120	0.63	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	1.26%#
50) Toluene-d8	10.058	98	9133	1.30	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	2.60%#
62) 4-Bromofluorobenzene	12.373	95	3418	1.35	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	2.70%#

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.836	85	1778	0.98	ug/l	89
3) Chloromethane	2.036	50	2527	0.95	ug/l	94
4) Vinyl Chloride	2.167	62	2606	0.97	ug/l	97
5) Bromomethane	2.547	94	2376	1.48	ug/l	98
6) Chloroethane	2.682	64	1450	0.91	ug/l #	68
7) Trichlorofluoromethane	2.997	101	2696	0.97	ug/l	93
8) Diethyl Ether	3.380	74	1307	0.97	ug/l	87
9) 1,1,2-Trichlorotrifluo...	3.714	101	1666	1.03	ug/l #	87
10) Methyl Iodide	3.904	142	1971	0.81	ug/l #	90
11) Tert butyl alcohol	4.756	59	682	1.86	ug/l	97
12) 1,1-Dichloroethene	3.695	96	2315	1.18	ug/l	98
13) Acrolein	3.573	56	1521	5.50	ug/l #	55
14) Allyl chloride	4.267	41	1652	0.44	ug/l #	51
15) Acrylonitrile	4.946	53	4779	4.67	ug/l #	60
16) Acetone	3.785	43	4622	5.41	ug/l	89
17) Carbon Disulfide	4.003	76	8121	1.40	ug/l #	94
18) Methyl Acetate	4.290	43	2439	1.02	ug/l #	83
19) Methyl tert-butyl Ether	4.997	73	3988	0.60	ug/l #	73
20) Methylene Chloride	4.502	84	2400	1.04	ug/l #	94
21) trans-1,2-Dichloroethene	4.984	96	2291	1.10	ug/l #	80
22) Diisopropyl ether	5.910	45	7026	1.00	ug/l #	87
23) Vinyl Acetate	5.852	43	26781	4.51	ug/l #	94
24) 1,1-Dichloroethane	5.798	63	3409	0.85	ug/l #	82
25) 2-Butanone	6.804	43	5376	4.38	ug/l	97
26) 2,2-Dichloropropane	6.769	77	2994	0.93	ug/l #	66
27) cis-1,2-Dichloroethene	6.781	96	2347	0.99	ug/l	98
28) Bromochloromethane	7.151	49	1722	1.06	ug/l #	78
29) Tetrahydrofuran	7.180	42	4975	5.19	ug/l #	81
30) Chloroform	7.328	83	3577	0.96	ug/l	98
31) Cyclohexane	7.611	56	7340	1.60	ug/l #	50
32) 1,1,1-Trichloroethane	7.527	97	2824	1.00	ug/l #	50
36) 1,1-Dichloropropene	7.749	75	2929	1.00	ug/l	96
37) Ethyl Acetate	6.888	43	2121	0.82	ug/l #	73
38) Carbon Tetrachloride	7.733	117	2154	0.88	ug/l	91
39) Methylcyclohexane	9.042	83	3790	1.07	ug/l #	85
40) Benzene	8.000	78	8327	0.92	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
 Data File : VN065574.D
 Acq On : 25 Jan 2021 13:13
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001

Quant Time: Jan 25 13:39:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:39:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.132	41	1681	1.19	ug/l #	100
42) 1,2-Dichloroethane	8.087	62	2224	0.80	ug/l #	74
43) Isopropyl Acetate	8.129	43	5282	1.03	ug/l #	81
44) Trichloroethene	8.801	130	2090	0.96	ug/l	97
45) 1,2-Dichloropropane	9.084	63	2187	0.88	ug/l	95
46) Dibromomethane	9.170	93	1226	0.90	ug/l	97
47) Bromodichloromethane	9.370	83	2755	0.93	ug/l #	91
48) Methyl methacrylate	9.167	41	2475	1.03	ug/l #	85
49) 1,4-Dioxane	9.177	88	667	17.62	ug/l #	93
51) 4-Methyl-2-Pentanone	9.952	43	13084	4.73	ug/l	96
52) Toluene	10.125	92	4897	0.93	ug/l	98
53) t-1,3-Dichloropropene	10.350	75	3014	0.88	ug/l	95
54) cis-1,3-Dichloropropene	9.807	75	3335	0.88	ug/l	98
55) 1,1,2-Trichloroethane	10.531	97	1926	0.95	ug/l	91
56) Ethyl methacrylate	10.399	69	3026	0.89	ug/l #	78
57) 1,3-Dichloropropane	10.675	76	3235	0.90	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.662	63	7204	4.08	ug/l	94
59) 2-Hexanone	10.727	43	8524	4.36	ug/l	99
60) Dibromochloromethane	10.871	129	1873	0.90	ug/l	86
61) 1,2-Dibromoethane	10.974	107	1790	0.89	ug/l	96
64) Tetrachloroethene	10.598	164	2278	1.12	ug/l	97
65) Chlorobenzene	11.402	112	4960	0.95	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.476	131	1711	0.94	ug/l #	65
67) Ethyl Benzene	11.482	91	9380	0.97	ug/l	95
68) m/p-Xylenes	11.592	106	6725	1.92	ug/l	92
69) o-Xylene	11.920	106	3244	0.95	ug/l	89
70) Styrene	11.936	104	4997	0.89	ug/l	93
71) Bromoform	12.100	173	1040	0.82	ug/l #	89
73) Isopropylbenzene	12.219	105	8257	1.03	ug/l	99
74) N-amyl acetate	12.048	43	3162	0.96	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	12.476	83	2541	1.08	ug/l #	95
76) 1,2,3-Trichloropropane	12.521	75	3371	1.36	ug/l #	100
77) Bromobenzene	12.498	156	1721	1.00	ug/l #	52
78) n-propylbenzene	12.563	91	10259	1.11	ug/l	94
79) 2-Chlorotoluene	12.646	91	6229	1.11	ug/l	89
80) 1,3,5-Trimethylbenzene	12.704	105	7393	1.10	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.273	75	747	0.89	ug/l	87
82) 4-Chlorotoluene	12.746	91	6413	1.15	ug/l	91
83) tert-Butylbenzene	12.964	119	6175	1.09	ug/l	92
84) 1,2,4-Trimethylbenzene	13.009	105	7274	1.09	ug/l	94
85) sec-Butylbenzene	13.141	105	8578	1.14	ug/l	97
86) p-Isopropyltoluene	13.260	119	7331	1.10	ug/l	94
87) 1,3-Dichlorobenzene	13.254	146	3278	1.11	ug/l	93
88) 1,4-Dichlorobenzene	13.254	146	3278	1.12	ug/l	94
89) n-Butylbenzene	13.588	91	6328	1.08	ug/l	91
90) Hexachloroethane	13.842	117	1127	0.96	ug/l	92
91) 1,2-Dichlorobenzene	13.624	146	3200	1.08	ug/l	92
92) 1,2-Dibromo-3-Chloropr...	14.244	75	355	0.83	ug/l	61
93) 1,2,4-Trichlorobenzene	14.884	180	1549	1.03	ug/l	100
94) Hexachlorobutadiene	14.980	225	1245	1.69	ug/l	94
95) Naphthalene	15.103	128	5395	1.10	ug/l #	92
96) 1,2,3-Trichlorobenzene	15.283	180	1894	1.25	ug/l	94


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
Data File : VN065574.D
Acq On : 25 Jan 2021 13:13
Operator : JC/MD
Sample : VSTDICC001
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Quant Time: Jan 25 13:39:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:39:20 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:168 Resp: 166385
Ion Ratio Lower Upper
168 100
99 59.7 42.6 63.8



#2
Dichlorodifluoromethane
Concen: 0.98 ug/l
RT: 1.836 min Scan# 30
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 85 Resp: 1778
Ion Ratio Lower Upper
85 100
87 26.8 16.6 49.8









#3
Chloromethane
Concen: 0.95 ug/l
RT: 2.036 min Scan# 92
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 50 Resp: 2527
Ion Ratio Lower Upper
50 100
52 35.5 25.8 38.8



#4
Vinyl Chloride
Concen: 0.97 ug/l
RT: 2.167 min Scan# 133
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 62 Resp: 2606
Ion Ratio Lower Upper
62 100
64 33.2 25.1 37.7









#5
Bromomethane
Concen: 1.48 ug/l
RT: 2.547 min Scan# 251
Delta R.T. 0.006 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 94 Resp: 2376
Ion Ratio Lower Upper
94 100
96 89.4 73.0 109.4



#6
Chloroethane
Concen: 0.91 ug/l
RT: 2.682 min Scan# 293
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 64 Resp: 1450
Ion Ratio Lower Upper
64 100
66 49.4 25.4 38.2#









#7
Trichlorofluoromethane
Concen: 0.97 ug/l
RT: 2.997 min Scan# 391
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:101 Resp: 2696
Ion Ratio Lower Upper
101 100
103 70.2 52.1 78.1



#8
Diethyl Ether
Concen: 0.97 ug/l
RT: 3.380 min Scan# 510
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 74 Resp: 1307
Ion Ratio Lower Upper
74 100
45 105.0 46.4 139.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.03 ug/l
RT: 3.714 min Scan# 614
Delta R.T. -0.006 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:101 Resp: 1666
Ion Ratio Lower Upper
101 100
85 56.1 34.6 51.8#
151 71.8 63.5 95.3



#10
Methyl Iodide
Concen: 0.81 ug/l
RT: 3.904 min Scan# 673
Delta R.T. -0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:142 Resp: 1971
Ion Ratio Lower Upper
142 100
127 34.6 31.7 47.5
141 8.0 10.9 16.3#









#11
Tert butyl alcohol
Concen: 1.86 ug/l
RT: 4.756 min Scan# 938
Delta R.T. -0.006 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampled :
VSTDIC001


Tgt Ion: 59 Resp: 682
Ion Ratio Lower Upper
59 100
57 8.8 8.0 12.0



#12
1,1-Dichloroethene
Concen: 1.18 ug/l
RT: 3.695 min Scan# 608
Delta R.T. -0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 96 Resp: 2315
Ion Ratio Lower Upper
96 100
61 155.7 122.5 183.7
98 64.6 50.8 76.2









#13
Acrolein
Concen: 5.50 ug/l
RT: 3.573 min Scan# 570
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 56 Resp: 1521
Ion Ratio Lower Upper
56 100
55 32.7 56.0 84.0#



#14
Allyl chloride
Concen: 0.44 ug/l
RT: 4.267 min Scan# 786
Delta R.T. -0.010 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 41 Resp: 1652
Ion Ratio Lower Upper
41 100
39 106.4 53.0 79.4#
76 62.6 28.9 43.3#









#15
Acrylonitrile
Concen: 4.67 ug/l
RT: 4.946 min Scan# 997
Delta R.T. 0.010 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 53 Resp: 4779
Ion Ratio Lower Upper
53 100
52 32.7 66.1 99.1#
51 32.5 27.9 41.9






#16
Acetone
Concen: 5.41 ug/l
RT: 3.785 min Scan# 636
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 43 Resp: 4622
Ion Ratio Lower Upper
43 100
58 40.2 27.0 40.6







Abundance Scan 705 (4.007 min): VN065570.D\data.ms (-68



#17

Carbon Disulfide
Concen: 1.40 ug/l
RT: 4.003 min Scan# 704
Delta R.T. -0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13


Tgt Ion: 76 Resp: 8121
Ion Ratio Lower Upper
76 100
78 11.8 7.8 11.6#




#18

Methyl Acetate
Concen: 1.02 ug/l
RT: 4.290 min Scan# 793
Delta R.T. 0.006 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 43 Resp: 2439
Ion Ratio Lower Upper
43 100
74 15.9 19.5 29.3#



Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001







#19

Methyl tert-butyl Ether
Concen: 0.60 ug/l
RT: 4.997 min Scan# 1013
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Instrument :
MSVOA_N
ClientSampled :
VSTDIC001




Tgt Ion: 73 Resp: 3988
Ion Ratio Lower Upper
73 100
57 34.4 17.3 25.9#




#20

Methylene Chloride
Concen: 1.04 ug/l
RT: 4.502 min Scan# 859
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 84 Resp: 2400
Ion Ratio Lower Upper
84 100
49 124.4 94.6 141.8
51 45.8 29.9 44.9#
86 65.0 51.4 77.2









#21
trans-1,2-Dichloroethene
Concen: 1.10 ug/l
RT: 4.984 min Scan# 1009
Delta R.T. -0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 96 Resp: 2291
Ion Ratio Lower Upper
96 100
61 108.9 104.9 157.3
98 48.4 51.8 77.8#



#22
Diisopropyl ether
Concen: 1.00 ug/l
RT: 5.910 min Scan# 1297
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 45 Resp: 7026
Ion Ratio Lower Upper
45 100
43 70.6 48.9 73.3
87 20.0 23.7 35.5#
59 9.7 9.9 14.9#









#23
 Vinyl Acetate
 Concen: 4.51 ug/l
 RT: 5.852 min Scan# 1279
 Delta R.T. 0.006 min
 Lab File: VN065574.D
 Acq: 25 Jan 2021 13:13




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 43 Resp: 26781
 Ion Ratio Lower Upper
 43 100
 86 8.4 8.6 13.0#



#24
 1,1-Dichloroethane
 Concen: 0.85 ug/l
 RT: 5.798 min Scan# 1262
 Delta R.T. -0.000 min
 Lab File: VN065574.D
 Acq: 25 Jan 2021 13:13

Tgt Ion: 63 Resp: 3409
 Ion Ratio Lower Upper
 63 100
 98 0.0 3.5 10.4#
 100 0.0 2.3 6.9#









#25
2-Butanone
Concen: 4.38 ug/l
RT: 6.804 min Scan# 1575
Delta R.T. 0.013 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampled :
VSTDIC001


Tgt Ion: 43 Resp: 5376
Ion Ratio Lower Upper
43 100
72 27.5 23.1 34.7



#26
2,2-Dichloropropane
Concen: 0.93 ug/l
RT: 6.769 min Scan# 1564
Delta R.T. -0.006 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 77 Resp: 2994
Ion Ratio Lower Upper
77 100
97 7.2 11.9 35.9#









#27
cis-1,2-Dichloroethene
Concen: 0.99 ug/l
RT: 6.781 min Scan# 1568
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 96 Resp: 2347




Ion	Ratio	Lower	Upper
96	100		
61	139.8	0.0	276.8
98	62.0	0.0	129.4




#28
Bromochloromethane
Concen: 1.06 ug/l
RT: 7.151 min Scan# 1683
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 49 Resp: 1722

Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.6
130	64.1	67.8	101.6#









#29
Tetrahydrofuran
Concen: 5.19 ug/l
RT: 7.180 min Scan# 1692
Delta R.T. 0.010 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 42 Resp: 4975
Ion Ratio Lower Upper
42 100
72 33.5 36.6 55.0#
71 30.9 34.3 51.5#



#30
Chloroform
Concen: 0.96 ug/l
RT: 7.328 min Scan# 1738
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 83 Resp: 3577
Ion Ratio Lower Upper
83 100
85 66.9 52.6 78.8









#31
Cyclohexane
Concen: 1.60 ug/l
RT: 7.611 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 56 Resp: 7340
Ion Ratio Lower Upper
56 100
69 103.4 26.6 39.8#
84 108.3 70.4 105.6#



#32
1,1,1-Trichloroethane
Concen: 1.00 ug/l
RT: 7.527 min Scan# 1800
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 97 Resp: 2824
Ion Ratio Lower Upper
97 100
99 0.0 50.8 76.2#
61 40.6 34.3 51.5









#33
1,2-Dichloroethane-d4
Concen: 1.43 ug/l
RT: 7.987 min Scan# 1943
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 65 Resp: 3120
Ion Ratio Lower Upper
65 100
67 60.4 0.0 107.2



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.547 min Scan# 2117
Delta R.T. -0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 114 Resp: 301352
Ion Ratio Lower Upper
114 100
63 22.1 0.0 38.6
88 17.6 0.0 30.6









#35
Dibromofluoromethane
Concen: 0.63 ug/l
RT: 7.543 min Scan# 1805
Delta R.T. -0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:113 Resp: 1120
Ion Ratio Lower Upper
113 100
111 194.6 82.2 123.2#
192 11.1 15.4 23.0#



#36
1,1-Dichloropropene
Concen: 1.00 ug/l
RT: 7.749 min Scan# 1869
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 75 Resp: 2929
Ion Ratio Lower Upper
75 100
110 33.0 18.2 54.6
77 30.1 24.9 37.3





#37

Ethyl Acetate

Concen: 0.82 ug/l

RT: 6.888 min Scan# 1601

Delta R.T. 0.003 min

Lab File: VN065574.D

Acq: 25 Jan 2021 13:13

Tgt Ion: 43 Resp: 2121

Ion Ratio Lower Upper

43 100

61 0.0 11.6 17.4#





70 5.0 8.6 13.0#

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001



#38

Carbon Tetrachloride

Concen: 0.88 ug/l

RT: 7.733 min Scan# 1864

Delta R.T. 0.003 min

Lab File: VN065574.D

Acq: 25 Jan 2021 13:13




Tgt Ion: 117 Resp: 2154


Ion Ratio Lower Upper

117 100

119 83.4 76.0 114.0

121 31.0 24.5 36.7









#39
Methylcyclohexane
Concen: 1.07 ug/l
RT: 9.042 min Scan# 2271
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	83	Resp:	3790
Ion	Ratio	Lower	Upper
83	100		
55	91.1	58.9	88.3#
98	48.1	35.9	53.9



#40
Benzene
Concen: 0.92 ug/l
RT: 8.000 min Scan# 1947
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:	78	Resp:	8327
Ion	Ratio	Lower	Upper
78	100		
77	24.1	19.0	28.4









#41
Methacrylonitrile
Concen: 1.19 ug/l
RT: 7.132 min Scan# 1677
Delta R.T. 0.006 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 41 Resp: 1681




Ion	Ratio	Lower	Upper
41	100		
39	35.7	0.0	0.0#
67	67.0	0.0	0.0#
52	12.3	0.0	0.0#




#42
1,2-Dichloroethane
Concen: 0.80 ug/l
RT: 8.087 min Scan# 1974
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 62 Resp: 2224

Ion	Ratio	Lower	Upper
62	100		
98	0.0	0.0	18.6









#43
Isopropyl Acetate
Concen: 1.03 ug/l
RT: 8.129 min Scan# 1987
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 5282
Ion Ratio Lower Upper
43 100
61 17.2 22.5 33.7#
87 7.3 11.1 16.7#



#44
Trichloroethene
Concen: 0.96 ug/l
RT: 8.801 min Scan# 2196
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 130 Resp: 2090
Ion Ratio Lower Upper
130 100
95 103.2 0.0 199.6









#45
1,2-Dichloropropane
Concen: 0.88 ug/l
RT: 9.084 min Scan# 2284
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 63 Resp: 2187
Ion Ratio Lower Upper
63 100
65 33.7 24.9 37.3



#46
Dibromomethane
Concen: 0.90 ug/l
RT: 9.170 min Scan# 2311
Delta R.T. -0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 93 Resp: 1226
Ion Ratio Lower Upper
93 100
95 81.1 67.0 100.6
174 101.5 83.2 124.8









#47
Bromodichloromethane
Concen: 0.93 ug/l
RT: 9.370 min Scan# 2373
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 83 Resp: 2755
Ion Ratio Lower Upper
83 100
85 58.1 51.0 76.4
127 0.0 7.0 10.6#



#48
Methyl methacrylate
Concen: 1.03 ug/l
RT: 9.167 min Scan# 2310
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 41 Resp: 2475
Ion Ratio Lower Upper
41 100
69 72.7 75.9 113.9#
39 57.4 46.6 69.8









#49
1,4-Dioxane
Concen: 17.62 ug/l
RT: 9.177 min Scan# 2313
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 88 Resp: 667
Ion Ratio Lower Upper
88 100
43 22.5 24.7 37.1#
58 63.3 52.6 78.8



#50
Toluene-d8
Concen: 1.30 ug/l
RT: 10.058 min Scan# 2587
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 98 Resp: 9133
Ion Ratio Lower Upper
98 100
100 64.5 50.5 75.7









#51
4-Methyl-2-Pentanone
Concen: 4.73 ug/l
RT: 9.952 min Scan# 2554
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 13084
Ion Ratio Lower Upper
43 100
58 38.1 32.5 48.7



#52
Toluene
Concen: 0.93 ug/l
RT: 10.125 min Scan# 2608
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 92 Resp: 4897
Ion Ratio Lower Upper
92 100
91 179.2 140.7 211.1









#53
t-1,3-Dichloropropene
Concen: 0.88 ug/l
RT: 10.350 min Scan# 2678
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 75 Resp: 3014
Ion Ratio Lower Upper
75 100
77 35.8 26.2 39.4



#54
cis-1,3-Dichloropropene
Concen: 0.88 ug/l
RT: 9.807 min Scan# 2509
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 75 Resp: 3335
Ion Ratio Lower Upper
75 100
77 30.2 24.9 37.3
39 47.0 36.2 54.4









#55
1,1,2-Trichloroethane
Concen: 0.95 ug/l
RT: 10.531 min Scan# 2734
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	97	Resp:	1926
Ion	Ratio	Lower	Upper
97	100		
83	95.5	67.7	101.5
85	51.4	44.5	66.7
99	67.3	49.4	74.0



#56
Ethyl methacrylate
Concen: 0.89 ug/l
RT: 10.399 min Scan# 2693
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:	69	Resp:	3026
Ion	Ratio	Lower	Upper
69	100		
41	79.3	49.6	74.4#
39	22.8	28.0	42.0#









#57
1,3-Dichloropropane
Concen: 0.90 ug/l
RT: 10.675 min Scan# 2779
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 76 Resp: 3235
Ion Ratio Lower Upper
76 100
78 32.4 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 4.08 ug/l
RT: 9.662 min Scan# 2464
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 63 Resp: 7204
Ion Ratio Lower Upper
63 100
106 24.9 22.6 33.8









#59
2-Hexanone
Concen: 4.36 ug/l
RT: 10.727 min Scan# 2795
Delta R.T. 0.006 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampled :
VSTDIC001


Tgt Ion: 43 Resp: 8524
Ion Ratio Lower Upper
43 100
58 55.5 28.1 84.3



#60
Dibromochloromethane
Concen: 0.90 ug/l
RT: 10.871 min Scan# 2840
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 129 Resp: 1873
Ion Ratio Lower Upper
129 100
127 64.5 38.5 115.5









#61
1,2-Dibromoethane
Concen: 0.89 ug/l
RT: 10.974 min Scan# 2872
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:107 Resp: 1790
Ion Ratio Lower Upper
107 100
109 90.8 75.6 113.4



#62
4-Bromofluorobenzene
Concen: 1.35 ug/l
RT: 12.373 min Scan# 3307
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 95 Resp: 3418
Ion Ratio Lower Upper
95 100
174 69.4 0.0 169.0
176 64.5 0.0 162.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.376 min Scan# 2997
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	117	Resp:	255297
Ion	Ratio	Lower	Upper
117	100		
82	58.1	43.7	65.5
119	32.2	26.1	39.1



#64
Tetrachloroethene
Concen: 1.12 ug/l
RT: 10.598 min Scan# 2755
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:	164	Resp:	2278
Ion	Ratio	Lower	Upper
164	100		
166	123.6	102.6	154.0
129	94.7	76.2	114.4
131	88.0	73.4	110.2









#65
Chlorobenzene
Concen: 0.95 ug/l
RT: 11.402 min Scan# 3005
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:112 Resp: 4960
Ion Ratio Lower Upper
112 100
114 31.8 26.1 39.1



#66
1,1,1,2-Tetrachloroethane
Concen: 0.94 ug/l
RT: 11.476 min Scan# 3028
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:131 Resp: 1711
Ion Ratio Lower Upper
131 100
133 96.7 48.0 144.0
119 0.0 34.1 102.2#









#67
Ethyl Benzene
Concen: 0.97 ug/l
RT: 11.482 min Scan# 3030
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 9380
Ion Ratio Lower Upper
91 100
106 29.2 25.5 38.3



#68
m/p-Xylenes
Concen: 1.92 ug/l
RT: 11.592 min Scan# 3064
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 106 Resp: 6725
Ion Ratio Lower Upper
106 100
91 213.9 161.4 242.0









#69
o-Xylene
Concen: 0.95 ug/l
RT: 11.920 min Scan# 3166
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:106 Resp: 3244
Ion Ratio Lower Upper
106 100
91 231.5 106.8 320.4



#70
Styrene
Concen: 0.89 ug/l
RT: 11.936 min Scan# 3171
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13


Tgt Ion:104 Resp: 4997
Ion Ratio Lower Upper
104 100
78 57.1 39.0 58.4
103 53.7 44.5 66.7








#71
Bromoform
Concen: 0.82 ug/l
RT: 12.100 min Scan# 3222
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001







Tgt Ion:173 Resp: 1040
Ion Ratio Lower Upper
173 100
175 40.1 23.8 71.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.315 min Scan# 3600
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:152 Resp: 93443
Ion Ratio Lower Upper
152 100
115 60.6 28.5 85.6
150 156.3 0.0 347.6









#73
Isopropylbenzene
Concen: 1.03 ug/l
RT: 12.219 min Scan# 3259
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:105 Resp: 8257
Ion Ratio Lower Upper
105 100
120 27.1 13.3 39.9



#74
N-ethyl acetate
Concen: 0.96 ug/l
RT: 12.048 min Scan# 3206
Delta R.T. 0.010 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 43 Resp: 3162
Ion Ratio Lower Upper
43 100
70 35.4 34.4 51.6
55 30.5 20.7 31.1
61 17.9 19.1 28.7#









#75
1,1,2,2-Tetrachloroethane
Concen: 1.08 ug/l
RT: 12.476 min Scan# 3339
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 83 Resp: 2541
Ion Ratio Lower Upper
83 100
131 4.4 5.3 15.8#
85 61.5 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 1.36 ug/l
RT: 12.521 min Scan# 3353
Delta R.T. -0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 75 Resp: 3371
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 1.00 ug/l
RT: 12.498 min Scan# 3346
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:156 Resp: 1721
Ion Ratio Lower Upper
156 100
77 294.2 96.3 288.8#
158 105.2 49.0 147.0



#78
n-propylbenzene
Concen: 1.11 ug/l
RT: 12.563 min Scan# 3366
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 91 Resp: 10259
Ion Ratio Lower Upper
91 100
120 20.6 11.8 35.3









#79
2-Chlorotoluene
Concen: 1.11 ug/l
RT: 12.646 min Scan# 3392
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 6229
Ion Ratio Lower Upper
91 100
126 28.4 17.4 52.3



#80
1,3,5-Trimethylbenzene
Concen: 1.10 ug/l
RT: 12.704 min Scan# 3410
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 105 Resp: 7393
Ion Ratio Lower Upper
105 100
120 46.8 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 0.89 ug/l
RT: 12.273 min Scan# 3276
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 75 Resp: 747
Ion Ratio Lower Upper
75 100
53 107.4 74.6 111.8
89 37.8 35.8 53.8



#82
4-Chlorotoluene
Concen: 1.15 ug/l
RT: 12.746 min Scan# 3423
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion: 91 Resp: 6413
Ion Ratio Lower Upper
91 100
126 29.4 17.2 51.6









#83
tert-Butylbenzene
Concen: 1.09 ug/l
RT: 12.964 min Scan# 3491
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	119	Resp:	6175
Ion	Ratio	Lower	Upper
119	100		
91	71.3	33.0	98.9
134	19.9	12.7	38.1



#84
1,2,4-Trimethylbenzene
Concen: 1.09 ug/l
RT: 13.009 min Scan# 3505
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:	105	Resp:	7274
Ion	Ratio	Lower	Upper
105	100		
120	48.5	22.4	67.0









#85
sec-Butylbenzene
Concen: 1.14 ug/l
RT: 13.141 min Scan# 3546
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:105 Resp: 8578
Ion Ratio Lower Upper
105 100
134 18.7 10.0 30.0



#86
p-Isopropyltoluene
Concen: 1.10 ug/l
RT: 13.260 min Scan# 3583
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:119 Resp: 7331
Ion Ratio Lower Upper
119 100
134 22.7 13.3 39.8
91 25.7 11.8 35.3









#87
1,3-Dichlorobenzene
Concen: 1.11 ug/l
RT: 13.254 min Scan# 3581
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	146	Resp:	3278
Ion	Ratio	Lower	Upper
146	100		
111	44.8	19.5	58.5
148	60.9	32.3	96.9



#88
1,4-Dichlorobenzene
Concen: 1.12 ug/l
RT: 13.254 min Scan# 3581
Delta R.T. -0.080 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:	146	Resp:	3278
Ion	Ratio	Lower	Upper
146	100		
111	44.8	19.0	57.0
148	60.9	31.8	95.3









#89
n-Butylbenzene
Concen: 1.08 ug/l
RT: 13.588 min Scan# 3685
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	91	Resp:	6328
Ion	Ratio	Lower	Upper
91	100		
92	45.8	26.1	78.3
134	21.7	13.0	38.9



#90
Hexachloroethane
Concen: 0.96 ug/l
RT: 13.842 min Scan# 3764
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:	117	Resp:	1127
Ion	Ratio	Lower	Upper
117	100		
201	68.8	37.8	113.3









#91
1,2-Dichlorobenzene
Concen: 1.08 ug/l
RT: 13.624 min Scan# 3696
Delta R.T. -0.000 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	146	Resp:	3200
Ion	Ratio	Lower	Upper
146	100		
111	46.3	20.1	60.3
148	58.5	32.0	96.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 0.83 ug/l
RT: 14.244 min Scan# 3889
Delta R.T. 0.006 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:	75	Resp:	355
Ion	Ratio	Lower	Upper
75	100		
155	48.5	44.1	132.4
157	72.7	56.0	168.2









#93
1,2,4-Trichlorobenzene
Concen: 1.03 ug/l
RT: 14.884 min Scan# 4088
Delta R.T. 0.006 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:180 Resp: 1549
Ion Ratio Lower Upper
180 100
182 94.8 47.3 142.0
145 31.2 15.6 46.8



#94
Hexachlorobutadiene
Concen: 1.69 ug/l
RT: 14.980 min Scan# 4118
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:225 Resp: 1245
Ion Ratio Lower Upper
225 100
223 69.8 31.9 95.9
227 60.1 31.6 95.0









#95
Naphthalene
Concen: 1.10 ug/l
RT: 15.103 min Scan# 4156
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion:128	Resp:	5395
Ion	Ratio	Lower Upper
128	100	
127	9.5	10.2 15.2#
129	8.0	8.6 13.0#



#96
1,2,3-Trichlorobenzene
Concen: 1.25 ug/l
RT: 15.283 min Scan# 4212
Delta R.T. 0.003 min
Lab File: VN065574.D
Acq: 25 Jan 2021 13:13

Tgt Ion:180	Resp:	1894
Ion	Ratio	Lower Upper
180	100	
182	89.2	47.4 142.3
145	29.7	16.5 49.5

