

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
 Data File : VN065576.D
 Acq On : 25 Jan 2021 14:43
 Operator : JC/MD
 Sample : VN0125MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0125MBL01

Quant Time: Jan 26 03:11:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	153098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	308618	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	311327	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	111006	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	105279	52.27	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.54%
35) Dibromofluoromethane	7.547	113	91515	50.04	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.08%
50) Toluene-d8	10.055	98	401073	55.79	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.58%
62) 4-Bromofluorobenzene	12.373	95	154581	59.71	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	119.42%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
Data File : VN065576.D
Acq On : 25 Jan 2021 14:43
Operator : JC/MD
Sample : VN0125MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0125MBL01

Quant Time: Jan 26 03:11:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

