

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
 Data File : VN065577.D
 Acq On : 25 Jan 2021 15:08
 Operator : JC/MD
 Sample : VN0125WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0125WBL01

Quant Time: Jan 26 01:45:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	202144	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	348065	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	298843	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	104270	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	127882	48.09	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.18%
35) Dibromofluoromethane	7.550	113	103414	50.14	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.28%
50) Toluene-d8	10.058	98	431982	53.28	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.56%
62) 4-Bromofluorobenzene	12.373	95	141199	48.36	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.72%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
Data File : VN065577.D
Acq On : 25 Jan 2021 15:08
Operator : JC/MD
Sample : VN0125WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0125WBL01

Quant Time: Jan 26 01:45:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

