

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
 Data File : VN065580.D
 Acq On : 25 Jan 2021 16:33
 Operator : JC/MD
 Sample : PB134285TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134285TB

Quant Time: Jan 26 01:47:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	205428	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	356767	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	307088	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	113783	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	130502	48.29	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.58%		
35) Dibromofluoromethane	7.550	113	105322	49.82	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.64%		
50) Toluene-d8	10.058	98	435975	52.46	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	104.92%		
62) 4-Bromofluorobenzene	12.373	95	147532	49.30	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	98.60%		
Target Compounds						
20) Methylene Chloride	4.502	84	12356	4.24	ug/l	95
43) Isopropyl Acetate	8.129	43	20079	3.29	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
Data File : VN065580.D
Acq On : 25 Jan 2021 16:33
Operator : JC/MD
Sample : PB134285TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134285TB

Quant Time: Jan 26 01:47:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

