

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
 Data File : VN065581.D
 Acq On : 25 Jan 2021 16:57
 Operator : JC/MD
 Sample : PB134285ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134285ZHE#01

Quant Time: Jan 26 01:48:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	199702	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	347613	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	303456	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	111659	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	127595	48.57	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.14%		
35) Dibromofluoromethane	7.550	113	100801	48.93	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.86%		
50) Toluene-d8	10.058	98	430149	53.12	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	106.24%		
62) 4-Bromofluorobenzene	12.373	95	144965	49.72	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	99.44%		
Target Compounds					Qvalue	
16) Acetone	3.779	43	6540	6.39	ug/l	99
20) Methylene Chloride	4.505	84	12595	4.45	ug/l	94
43) Isopropyl Acetate	8.129	43	19086	3.21	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
Data File : VN065581.D
Acq On : 25 Jan 2021 16:57
Operator : JC/MD
Sample : PB134285ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134285ZHE#01

Quant Time: Jan 26 01:48:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

