

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
 Data File : VN065583.D
 Acq On : 25 Jan 2021 17:46
 Operator : JC/MD
 Sample : PB134285ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134285ZHE#03

Quant Time: Jan 26 01:48:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	201838	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	352025	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	307687	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	112037	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	130389	49.10	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.20%
35) Dibromofluoromethane	7.547	113	102814	49.29	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.58%
50) Toluene-d8	10.058	98	438083	53.42	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.84%
62) 4-Bromofluorobenzene	12.373	95	146564	49.64	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.28%
Target Compounds						
					Qvalue	
16) Acetone	3.782	43	6196	5.99	ug/l	89
20) Methylene Chloride	4.505	84	13313	4.65	ug/l	96
43) Isopropyl Acetate	8.129	43	20362	3.38	ug/l #	88
51) 4-Methyl-2-Pentanone	9.952	43	3696	1.11	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
Data File : VN065583.D
Acq On : 25 Jan 2021 17:46
Operator : JC/MD
Sample : PB134285ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134285ZHE#03

Quant Time: Jan 26 01:48:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

