

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
 Data File : VN065588.D
 Acq On : 25 Jan 2021 19:48
 Operator : JC/MD
 Sample : M1218-04 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 238

Quant Time: Jan 26 04:14:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	206926	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	356634	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	299478	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	104351	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	129985	47.75	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.50%		
35) Dibromofluoromethane	7.547	113	103560	49.00	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.00%		
50) Toluene-d8	10.058	98	436065	52.49	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	104.98%		
62) 4-Bromofluorobenzene	12.373	95	139391	46.60	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.20%		
Target Compounds						Qvalue
16) Acetone	3.785	43	4681	4.42	ug/l	97
40) Benzene	7.997	78	11281	1.03	ug/l	98
52) Toluene	10.122	92	26621	4.10	ug/l	98
67) Ethyl Benzene	11.482	91	3662	0.31	ug/l	98
68) m/p-Xylenes	11.588	106	9795	2.30	ug/l	94
69) o-Xylene	11.920	106	6814	1.65	ug/l	97
80) 1,3,5-Trimethylbenzene	12.701	105	4213	0.54	ug/l	97
84) 1,2,4-Trimethylbenzene	13.010	105	14993	1.93	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012521\
Data File : VN065588.D
Acq On : 25 Jan 2021 19:48
Operator : JC/MD
Sample : M1218-04 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
238

Quant Time: Jan 26 04:14:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

