

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012522\
 Data File : VN070768.D
 Acq On : 25 Jan 2022 11:54
 Operator : JC/MD
 Sample : VN0125MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0125MBL01

Quant Time: Jan 26 00:47:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	898128	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1430533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1237696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	426573	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	617712	47.990	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.980%
35) Dibromofluoromethane	8.024	113	424979	47.892	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.780%
50) Toluene-d8	10.440	98	1757748	47.021	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.040%
62) 4-Bromofluorobenzene	12.731	95	604991	45.416	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012522\
Data File : VN070768.D
Acq On : 25 Jan 2022 11:54
Operator : JC/MD
Sample : VN0125MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0125MBL01

Quant Time: Jan 26 00:47:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

