

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012522\
 Data File : VN070772.D
 Acq On : 25 Jan 2022 14:23
 Operator : JC/MD
 Sample : N1250-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3

Quant Time: Jan 26 00:48:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	893532	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1416815	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1256981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	487041	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	607776	47.461	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.920%		
35) Dibromofluoromethane	8.021	113	417455	47.500	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.000%		
50) Toluene-d8	10.440	98	1741858	47.047	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	94.100%		
62) 4-Bromofluorobenzene	12.731	95	646791	49.024	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	98.040%		
Target Compounds						
					Qvalue	
16) Acetone	4.288	43	46815	5.267	ug/l	92
39) Methylcyclohexane	9.459	83	10360	0.567	ug/l	91
40) Benzene	8.458	78	41803	0.912	ug/l	95
59) 2-Hexanone	11.087	43	18135	1.304	ug/l	90
67) Ethyl Benzene	11.843	91	24515	0.486	ug/l	95
68) m/p-Xylenes	11.953	106	6525	0.355	ug/l	92
69) o-Xylene	12.283	106	3658	0.201	ug/l	68
73) Isopropylbenzene	12.575	105	33287	0.722	ug/l	99
78) n-propylbenzene	12.918	91	63734	1.237	ug/l	95
84) 1,2,4-Trimethylbenzene	13.366	105	48108	1.343	ug/l	95
85) sec-Butylbenzene	13.500	105	27730	0.629	ug/l	96
86) p-Isopropyltoluene	13.613	119	12974	0.375	ug/l	98
89) n-Butylbenzene	13.940	91	21339	0.747	ug/l #	82
91) 1,2-Dichlorobenzene	13.986	146	10638	0.627	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012522\
Data File : VN070772.D
Acq On : 25 Jan 2022 14:23
Operator : JC/MD
Sample : N1250-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3

Quant Time: Jan 26 00:48:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

