

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012522\
 Data File : VN070773.D
 Acq On : 25 Jan 2022 14:48
 Operator : JC/MD
 Sample : N1250-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Jan 26 00:48:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	866545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1382926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1225085	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	438651	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	586599	47.234	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.460%
35) Dibromofluoromethane	8.021	113	409398	47.725	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.440%
50) Toluene-d8	10.440	98	1710820	47.341	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.680%
62) 4-Bromofluorobenzene	12.731	95	595577	46.249	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.500%
Target Compounds						
16) Acetone	4.296	43	19183	2.225	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012522\
Data File : VN070773.D
Acq On : 25 Jan 2022 14:48
Operator : JC/MD
Sample : N1250-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Jan 26 00:48:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

