

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012523\
 Data File : VN076391.D
 Acq On : 25 Jan 2023 10:47
 Operator : JC\MD
 Sample : VN0125MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0125MBL01

Quant Time: Jan 26 07:13:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	339771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	545791	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	459842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	171641	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	201096	52.859	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.720%
35) Dibromofluoromethane	8.177	113	169103	49.530	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.060%
50) Toluene-d8	10.571	98	630586	50.011	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.020%
62) 4-Bromofluorobenzene	12.853	95	185504	45.067	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012523\
Data File : VN076391.D
Acq On : 25 Jan 2023 10:47
Operator : JC\MD
Sample : VN0125MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0125MBL01

Quant Time: Jan 26 07:13:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

