

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012523\
 Data File : VN076392.D
 Acq On : 25 Jan 2023 11:11
 Operator : JC\MD
 Sample : VN0125WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0125WBL01

Quant Time: Jan 26 07:13:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	366779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	591266	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	533550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	216445	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	217768	53.027	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.060%
35) Dibromofluoromethane	8.171	113	185026	50.026	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.060%
50) Toluene-d8	10.571	98	694527	50.846	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.700%
62) 4-Bromofluorobenzene	12.853	95	220647	49.482	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012523\
Data File : VN076392.D
Acq On : 25 Jan 2023 11:11
Operator : JC\MD
Sample : VN0125WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0125WBL01

Quant Time: Jan 26 07:13:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

