

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012523\
 Data File : VN076397.D
 Acq On : 25 Jan 2023 13:22
 Operator : JC\MD
 Sample : 01291-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW10-GW-358-360

Quant Time: Jan 26 07:16:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	388273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	630105	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	537995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	206430	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	223677	51.450	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.900%
35) Dibromofluoromethane	8.171	113	192497	48.838	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.680%
50) Toluene-d8	10.571	98	722237	49.616	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.240%
62) 4-Bromofluorobenzene	12.853	95	222018	46.720	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012523\
Data File : VN076397.D
Acq On : 25 Jan 2023 13:22
Operator : JC\MD
Sample : 01291-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW10-GW-358-360

Quant Time: Jan 26 07:16:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

