

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012523\
 Data File : VN076403.D
 Acq On : 25 Jan 2023 15:46
 Operator : JC\MD
 Sample : 01291-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW10-TB-20230120

Quant Time: Jan 26 07:17:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	450767	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	730743	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	641777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	240063	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	275959	54.676	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.360%	
35) Dibromofluoromethane	8.171	113	225709	49.377	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.760%	
50) Toluene-d8	10.571	98	856849	50.756	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.520%	
62) 4-Bromofluorobenzene	12.853	95	272850	49.509	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.020%	
Target Compounds						
20) Methylene Chloride	5.289	84	10667	1.645	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012523\
Data File : VN076403.D
Acq On : 25 Jan 2023 15:46
Operator : JC\MD
Sample : 01291-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW10-TB-20230120

Quant Time: Jan 26 07:17:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

