

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012523\
 Data File : VN076401.D
 Acq On : 25 Jan 2023 14:58
 Operator : JC\MD
 Sample : 01291-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW10-DUP-20230124

Quant Time: Jan 26 07:17:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	377878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	619065	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	522820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	192699	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	223422	52.805	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.620%
35) Dibromofluoromethane	8.171	113	191752	49.516	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.040%
50) Toluene-d8	10.571	98	711504	49.750	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.500%
62) 4-Bromofluorobenzene	12.853	95	210513	45.089	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.180%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012523\
Data File : VN076401.D
Acq On : 25 Jan 2023 14:58
Operator : JC\MD
Sample : 01291-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW10-DUP-20230124

Quant Time: Jan 26 07:17:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1050
Delta R.T. -0.000 min
Lab File: VN076401.D
Acq: 25 Jan 2023 14:58




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW10-DUP-20230124


Tgt Ion:168 Resp: 377878
Ion Ratio Lower Upper
168 100
99 58.0 41.6 62.4



#33
1,2-Dichloroethane-d4
Concen: 52.805 ug/l
RT: 8.583 min Scan# 1110
Delta R.T. 0.001 min
Lab File: VN076401.D
Acq: 25 Jan 2023 14:58

Tgt Ion: 65 Resp: 223422
Ion Ratio Lower Upper
65 100
67 50.7 0.0 104.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1199

Delta R.T. 0.000 min

Lab File: VN076401.D

Acq: 25 Jan 2023 14:58

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-RW10-DUP-20230124





Tgt Ion:114 Resp: 619065

Ion Ratio Lower Upper

114 100

63 18.1 0.0 35.8

88 15.4 0.0 28.0



#35

Dibromofluoromethane

Concen: 49.516 ug/l

RT: 8.171 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VN076401.D

Acq: 25 Jan 2023 14:58




Tgt Ion:113 Resp: 191752


Ion Ratio Lower Upper

113 100

111 101.5 82.7 124.1

192 18.6 15.6 23.4









#50
Toluene-d8
Concen: 49.750 ug/l
RT: 10.571 min Scan# 1448
Delta R.T. -0.000 min
Lab File: VN076401.D
Acq: 25 Jan 2023 14:58




Instrument : MSVOA_N
ClientSampleId : BP-VPB-RW10-DUP-20230124


Tgt Ion: 98 Resp: 711504
Ion Ratio Lower Upper
98 100
100 65.0 52.2 78.4



#62
4-Bromofluorobenzene
Concen: 45.089 ug/l
RT: 12.853 min Scan# 1836
Delta R.T. 0.006 min
Lab File: VN076401.D
Acq: 25 Jan 2023 14:58

Tgt Ion: 95 Resp: 210513
Ion Ratio Lower Upper
95 100
174 83.5 0.0 174.6
176 79.7 0.0 167.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN076401.D
Acq: 25 Jan 2023 14:58

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW10-DUP-20230124

Tgt Ion:117 Resp: 522820
Ion Ratio Lower Upper
117 100
82 52.0 42.2 63.2
119 30.8 26.9 40.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 1996
Delta R.T. -0.000 min
Lab File: VN076401.D
Acq: 25 Jan 2023 14:58

Tgt Ion:152 Resp: 192699
Ion Ratio Lower Upper
152 100
115 58.2 29.6 88.9
150 157.5 0.0 351.4

